



SAFETY AND BUILDINGS DIVISION  
Plumbing Product Review  
P.O. Box 2658  
Madison, Wisconsin 53701-2658  
**TTY: Contact Through Relay**

**Scott Walker, Governor**  
**Dave Ross, Secretary**

May 15, 2012

STERIS CORPORATION  
JASON BEAN  
5960 HEISLEY RD.  
MENTOR OH 44060

Re: Description: HEALTH CARE PLUMBING APPLIANCE  
Manufacturer: STERIS CORPORATION  
Product Name: SYSTEM 1E LIQUID CHEMICAL STERILANT PROCESSING SYSTEM  
Model Number(s): P6500  
Product File No: 20120178

The specifications and/or plans for this plumbing product have been reviewed and determined to be in compliance with chapters SPS 382 through 384, Wisconsin Administrative Code, and Chapters 145 and 160, Wisconsin Statutes.

The Department hereby issues an approval based on the Wisconsin Statutes and the Wisconsin Administrative Code. This approval is valid until the end of May 2017.

This approval is contingent upon compliance with the following stipulation(s):

- The indirect waste pipe from this product must discharge by means of an air-gap, an air-gap fitting that conforms to ASME A112.1.3 or an air-gap fitting approved by the Department.
- A reduced pressure principle backflow preventer (ASSE 1013 or CAN/CSA B64.4) or an assembly or method suitable for high hazard must be installed on the water supply pipes serving this product.

The department is in no way endorsing this product or any advertising, and is not responsible for any situation which may result from its use.

Sincerely,

Jerry Thompson  
Plumbing Product Reviewer  
Phone: 608-266-6742  
Fax: 608-267-9566  
E-mail: jerry.thompson@wi.gov